



# NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,  
Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,  
Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370  
E-mail : [ncqc@calibrationlaboratory.in](mailto:ncqc@calibrationlaboratory.in), • [calibrationlab.ncqc@gmail.com](mailto:calibrationlab.ncqc@gmail.com)



Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in) • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

## Calibration Certificate

**TRUE COPY**

F/CR/E/024/03 Issue No.02

### Name of Customer

NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS,  
Iskcon - Ambli Road, Ambli, Ahmedabad - 380 058,  
Gujarat

### Certificate No.

NCQC-E/010425/003

### Date of Issue

12/04/2025

### Date of Calibration

01/04/2025 to 07/04/2025

### Suggested Due Date

06/04/2026

Date of Receipt / Ref.No.: 01/04/2025

ULR No. : CC212825000000949F

Discipline : Electro-Technical Calibration  
Time & Frequency

### Details of Observation of Unit Under Calibration

Identification No. : NCQC/DTSC-02

Serial No. : =====

Name of Instrument : Digital Timer/Stop watch Calibrator

### Range

0 to 99999.999 Sec

### Resolution

0.001 Sec

### Make / Model

Digital Instrumentation/=====

### Location

Laboratory

### Initial Error

Nil

### Visual Inspection

OK

### Accuracy

As per Manual

### Type / Size

Digital

### Loc. of Calibration

At Laboratory

### Environment Cond.

25 ± 5°C, 30 to 75% Rh.

### Actual Env. Cond.

25.3 °C 50.8 %RH

### Std. Ref. Used

NIST 960-12

### Cali.Method No.

NCQC/CM/E/024

### Master Instruments Detail

NCQC SYS.NO.	Nomenclature	Make/ Model	Id / Serial No.	Traceable To	Certificate No.	Valid Upto
374	Timer / Stop watch Calibrator	Digital Instrumentation / =====	NCQC/E-29 / DITC080519	CC-4062	CC40622500000036 8F	06/03/2026

### Calibration Results

Set Points on UUC	Unit	Reading observed by Reference in seconds	Absolute Error in second	Expanded Uncertainty in Second (k =2) ±
0.010	Sec	0.010	0.000	0.002
60.000	Sec	60.003	0.003	0.036
900.000	Sec	900.004	0.004	0.116
1800.000	Sec	1800.008	0.008	0.221
3600.000	Sec	3600.042	0.042	0.443
43200.000	Sec	43200.711	0.711	5.464
86390.000	Sec	86392.393	2.393	23.250

NCQC System Certificate No. 424

Traceable To National / International Standards.

### Calibrated By

Krupesh Gajera  
Calibration Engineer

### Reviewed & Approved By



NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"





# NCQC LABORATORY LLP

4, Abhishree Corporate Park, Nr. Swagat Bunglow BRTS,  
Iskcon-Ambli Road, Ambli, Ahmedabad-380 058 • Ph. +91-79-29795322, 29795323,  
Fax : +91-79-29795323, Cell No. +91-9327017517, +91-9328616370  
E-mail : [ncqc@calibrationlaboratory.in](mailto:ncqc@calibrationlaboratory.in), • [calibrationlab.ncqc@gmail.com](mailto:calibrationlab.ncqc@gmail.com)



Visit our Web Site : [www.calibrationlaboratory.in](http://www.calibrationlaboratory.in) • LLP IN : ACH-9094 / GSTIN : 24AAWFN8511B1Z9

Precision Calibration Centre with National / International Traceability for Temperature, Dimensional, Pressure, Vacuum, Time, Mass, Electrical, Noise, Airflow, Lux & all Special Purpose Instruments in all ranges.

Calibration Certificate		F/CR/E/024/03 Issue No.02	
<b>Name of Customer</b> NCQC LABORATORY LLP 4, Abhishree Corporate Park, Nr. Swagat Bungalows BRTS, Iskcon – Ambli Road, Ambli, Ahmedabad – 380 058, Gujarat		<b>Certificate No.</b> NCQC-E/010425/003	
		<b>Date of Issue</b> 12/04/2025	
		<b>Date of Calibration</b> 01/04/2025 to 07/04/2025	
		<b>Suggested Due Date</b> 06/04/2026	
		<b>Date of Receipt / Ref.No.:</b> 01/04/2025	
<b>ULR No. :</b> CC21282500000949F		<b>Discipline :</b> Electro-Technical Calibration Time & Frequency	
<b>Details of Observation of Unit Under Calibration</b>	<b>Identification No. :</b> NCQC/DTSC-02		
	<b>Serial No. :</b> =====		
	<b>Name of Instrument :</b> Digital Timer/Stop watch Calibrator		
<b>Remarks:</b> <ul style="list-style-type: none"><li>• Averages of minimum three readings are reported.</li><li>• Suggested due date is given based on customer requirements.</li><li>• These results are obtained at the time of calibration.</li><li>• Condition of instrument found satisfactory during receipt</li><li>• Result relates to the item calibrated only.</li><li>• No external provider was used for calibration and hence it is not applicable.</li><li>• Calibration certificate shall not be reproduced except in full without written approval of Director, NCQC.</li><li>• Any hand written corrections (except @) or photocopies of the report invalidates this certificate.</li><li>• The uncertainties are for a confidence probability of not less than 95% with coverage factor k = 2</li><li>• Our masters are directly calibrated through NABL accredited calibration laboratory having direct traceability with National / international standard.</li></ul>			
Traceable To National / International Standards.			
<b>Calibrated By</b>  Krupesh Gajera Calibration Engineer	<b>Reviewed &amp; Approved By</b>  Anchal Calibration Incharge		

\*\*\* End of Certificate \*\*\*



NCQC DEFINES CALIBRATION AS "PRECISION AND RELIABILITY OF INSTRUMENTS FOR YOUR BETTER TOMORROW"